



(585) 359-0130 • Fax: (585) 359-0167
6789 West Henrietta Road • Rush, NY 14543

CERTIFICATE OF COMPLIANCE

Date: 11/16/2020

Lot Number: 2010

Expiration: 2023/10

The applicable technical information/specification for the product below has been met:

BDS-050 AirView™ Bowie-Dick Type Test Sheet. Used to detect the presence of residual air in a pre-vacuum steam sterilizer operating at 134°C. The BD Test Sheet used in a standard AAMI towel pack is a reliable tool for monitoring air removal in steam pre-vacuum processes.

We certify that the supplied consignment of the commodity described above was produced by or under the direction of Crosstex/SPSmedical and conforms to specification requirements including special processes, if any, and that physical, biological and sterility test reports or process certification are on file.

Specific Standards and Conditions:

Process at 134°C (273°F) for 3.5-4 minutes. Do not exceed 4 minutes. The indicator test sheet shall show a uniform color change.

Darcy Darling

By: _____
Quality Assurance