



(585) 359-0130 • Fax: (585) 359-0167  
6789 West Henrietta Road • Rush, NY 14543

## CERTIFICATE OF COMPLIANCE

Date: 12/02/2021

Lot Number: 2108

Expiration: 2024/08/31

The applicable technical information/specification for the product below has been met:

**BDS-050** AirView™ Bowie-Dick Type Test Sheet. Used to detect the presence of residual air in a pre-vacuum steam sterilizer operating at 134°C. The BD Test Sheet used in a standard AAMI towel pack is a reliable tool for monitoring air removal in steam pre-vacuum processes.

We certify that the supplied consignment of the commodity described above was produced by or under the direction of Crosstex/SPSmedical and conforms to specification requirements including special processes, if any, and that physical, biological and sterility test reports or process certification are on file.

### Specific Standards and Conditions:

Process at 134°C (273°F) for 3.5-4 minutes. Do not exceed 4 minutes. The indicator test sheet shall show a uniform color change.

*Darcy Darling*

By: \_\_\_\_\_  
Quality Assurance