

CERTIFICATE OF ANALYSIS


This lot of product meets, where applicable, the accepted performance criteria recommended in ISO 11140-1 for Class 1 compliance.

PRODUCT NAME: Chemical Process Indicator


PRODUCT CODE: CPI-R03

LOT: 19R0516-01-01

For use in Monitoring: Radiation (Gamma and E-beam)

 16-May-2022

Origin of Manufacture: United States






 16-May-2019

Quantity: 5000/Roll

Characteristic	Specification	Result
Color Transition (≥ 10 kGy)	Yellow to Red	Pass
Date of Performance Verification:	August 2019	

Color Transition: Color transition may vary. A color transition from the initial color to a shade of the signal color is considered an acceptable transition and verifies exposure to the sterilization process.

Storage:

	15°C to 30°C		Protect from heat and radioactive sources
	20% to 70% relative humidity		Keep dry
	Do not use damaged Indicators or Indicators which have transitioned to red. Do not use after expiration date		

Bryan Case 11-13-2020

QUALITY APPROVAL