

CERTIFICATE OF ANALYSIS


This lot of product meets, where applicable, the accepted performance criteria recommended in ISO 11140-1 for Type 1 compliance.

PRODUCT NAME: Chemical Process Indicator


PRODUCT CODE: CPI-E03

LOT : 20E0528

For use in Monitoring: Ethylene Oxide

 : 28-MAY-2022

Origin of Manufacture: United States





 : 28-MAY-2020

Quantity: 5,000/ Roll

Characteristic	Specification	Result
Color Transition (20 minutes at 600 mg/L, 54°C and 60% RH)	Purple to Green	Pass
Date of Performance Verification	May 2020	

Color Transition: Color transition may vary. A color transition from the initial color to a shade of the signal color is considered an acceptable transition and verifies exposure to the sterilization process.

Storage:

	15°C to 30°C		Keep dry
	20% to 70% relative humidity		Keep away from sterilants. Do not use after expiration date

Darcy Darling

QUALITY APPROVAL