

## CERTIFICATE OF ANALYSIS

This lot of product meets, where applicable, the accepted performance criteria recommended in ISO 11140-1 for Type 1 compliance.

**PRODUCT NAME:** Chemical Process Indicator

**PRODUCT CODE:** CPI-E03

**LOT**

: 20E0817

For use in  
Monitoring: Ethylene Oxide



: 17-AUG-2022

Origin of  
Manufacture: United States



: 17-AUG-2020

Quantity: 5000/ROLL

Characteristic	Specification	Result
Color Transition (20 minutes at 600 mg/L, 54°C and 60% RH)	Purple to Green	Pass
Date of Performance Verification	September 2020	

**Color Transition:** Color transition may vary. A color transition from the initial color to a shade of the signal color is considered an acceptable transition and verifies exposure to the sterilization process.

**Storage:**

	15°C to 30°C		Keep dry
	20% to 70% relative humidity		Keep away from sterilants. Do not use after expiration date

*Darcy Darling*

10/20/2020

QUALITY APPROVAL