

Certificate of Analysis



This lot of product meets the accepted performance criteria recommended in ISO 11140-1: 2014 as a Type 1 Process Indicator.

Product Name: Chemical Process Indicator for Monitoring Irradiation Processes

For use in Monitoring: Irradiation (Gamma)

Appearance: Yellow to Red, 12 mm Imprinted Dot

LOT R210922

REF CRYR-3C

22-SEP-2024

22-SEP-2021

Characteristic	Specification	Result
Radiation Cobalt ⁶⁰ , 10 kGy	Yellow to Red Color Transition	Pass
Date of Product Performance Verification	October 2021	

Storage and Shelf Life			
	Room Temperature (15° - 30°C)		Keep away from sunlight
	20% to 80% Relative Humidity		Protect from heat, radioactive sources, and sterilizing agents
Shelf Life	36 Months from the date of manufacture		

Disposal: Discard as general waste.

[Signature]

Quality Approval

Date

July 1, 2022

True Indicating, LLC
 946 Kane Street, Suite A
 Toledo, Ohio 43612
 www.trueindicating.com
 info@trueindicating.com
 FOR LAB/INDUSTRIAL USE ONLY.