



(585) 359-0130 • Fax: (585) 359-0167
6789 West Henrietta Road • Rush, NY 14543

CERTIFICATE OF COMPLIANCE

Date: 06/21/2022

Lot Number: 2205

Expiration Date: 2024-05-31

The applicable technical information/specification for the product below has been met:

GPS-250R 4-inch Hydrogen Peroxide indicator strips printed with proprietary SPSmedical indicator ink. Strips for use in the STERRAD 100S, 200, NX, and Steris V-Pro sterilization cycles.

We certify that the supplied consignment of the commodity described above was produced by or under the direction of SPSmedical and conforms to specification requirements including special processes, if any, and that physical, biological and sterility test reports or process certification are on file.

Specific Standards and Conditions:

The hydrogen peroxide indicator ink shall turn Blue when processed in full STERRAD® 100S, 200, NX, 100NX and Steris V-PRO sterilization cycles.

Darcy Darling

By: _____
Quality Assurance