



(585) 359-0130 • Fax: (585) 359-0167 6789 West Henrietta Road • Rush, NY 14543

## **CERTIFICATE OF COMPLIANCE**

Date: <u>10/27/2023</u>

Lot Number: 613

Expiration: 2025-11-01

The applicable technical information/specification for the product below has been met:

SSI-250 Steam*Plus*<sup>™</sup> Steam Sterilization Integrators conforms to AAMI/ANSI ST-60, EN 867 and suggested performance parameters of ISO 11140-1:2005/EN ISO 11140-1:2009 (class 5). Compliance to ISO 11140 (class 5) allows the product to be used at any temperature between 121C and 135C.

We certify that the supplied consignment of the commodity described above was produced by or under the direction of SPSmedical and conforms to specification requirements including special processes, if any, and that physical, biological and sterility test reports or process certification if required, are on file.

## **Specific Standards and Conditions:**

The Steam*Plus*<sup>™</sup> Steam Sterilization Integrators when processed in a steam sterilizer at 250°F for 30 minutes or 270°F for 3 minutes, the blue bar will be visible in the "SAFE" area. Product should be stored in a cool dry place.

Darcy Darling

By:\_\_\_\_\_