





CERTIFICATE OF CONFORMANCE






Product:	GPS-250R	Description:	VH2O2 Chemical Indicator Strips
LOT :	2302	Quantity:	250 strips/box
 :	2025-02-28	Origin of Manufacture:	United States
 :	2023-02-28	For use in Monitoring:	Hydrogen Peroxide
ISO 11140-1 Type 1 Compliant			

Specific Standards and Conditions:

The hydrogen peroxide indicator ink shall turn Blue when processed in full cycles within the STERRAD® 100S (Standard & Long), 200, 100NX® (Standard, Duo, Flex & Express), NX® (Standard & Advanced), the STERIS® V-PRO® 1, V-PRO® 1 Plus (Lumen & Non-lumen), V-PRO® maX (Flexible, Lumen & Non-lumen), and the Sterilucent™ PSD-85 (Lumen & Non-lumen).

We certify that the supplied consignment of the commodity described above was produced by or under the direction of Crosstex and conforms to specification requirements including special processes, if any, and that physical, biological and sterility test reports or process certification are on file.

Storage and Shelf Life:

 59°F/15°C	86°F/30°C	15°C to 30°C (59°F to 86°F)		Keep away from light sources
 30%	70%	30% to 70% Relative Humidity		Do not re-use
	Do not freeze. Keep away from sterilants and excessive heat. Do not use beyond the expiration date provided on the product packaging.			

Darcy Darling

04/17/2023

QUALITY ASSURANCE

Date



Crosstex International, Inc.
6789 W. Henrietta Road, Rush, NY 14543 USA
585.359.0130 | 631.582.6777 | crosstex.com

All company and product names are trademarks of Hu-Friedy Mfg. Co., LLC, its affiliates or related companies, unless otherwise noted.

All other product and company names referenced are trademarks of their respective owner.

©2022 Hu-Friedy Mfg. Co., LLC. All rights reserved.

CERT-00057/Rev A - 10/2022