

Certificate of Analysis

This lot of product meets the accepted performance criteria recommended in ISO 11140-1: 2014 as a Type 5 Integrating Indicator.

Product Name: Indicator for Monitoring Ethylene Oxide (EO) Processes

For use in Monitoring: Ethylene Oxide (EO)

Appearance: EO Indicator Strip Label, 19 mm x 64 mm

LOT

E231130

REF

CEVG-15C

M

30-NOV-2023

23

30-NOV-2026

Characteristic	Specification	Result	
Ethylene Oxide (100%) 54°C, 40 minutes, 600 mg/L, 60% RH and 37°C, 75 minutes, 600 mg/L, 60% RH	Violet to Green Color Transition	Pass	
Date of Product Performance Verification:	December 2023		

Storage and Shelf Life				
+15°C	Room Temperature (15° – 30°C)	誉	Keep away from sunlight	
20%	20% to 80% Relative Humidity	***	Protect from heat, radioactive sources, and sterilizing agents.	
Shelf Life	36 Months from the date of manufact	ıre		

Disposal: Discard as general waste.

Quality Approval December 15, 2023

Date

True Indicating, LLC 946 Kane Street, Suite A Toledo, Ohio 43612 www.trueindicating.com

info@trueindicating.com